



FM Approvals  
1151 Boston-Providence Turnpike  
P.O. Box 9102 Norwood, MA 02062 USA  
T: 781 762 4300 F: 781 762 9375 www.fmglobal.com

# CERTIFICATE OF COMPLIANCE

## HAZARDOUS (CLASSIFIED) LOCATION ELECTRICAL EQUIPMENT

This certificate is issued for the following equipment:

**3000MB. Paramagnetic Oxygen Analyzer.**  
NI / I / 2 / ABCD / T3 Ta = 45°C

Equipment Ratings:

Nonincendive for use in Class I, Division 2, Groups A, B, C and D Hazardous (Classified) Locations

Approved for:

Teledyne Analytical Instruments  
16830 Chestnut St.  
City of Industry, CA 91749



This certifies that the equipment described has been found to comply with the following FM Approval Standards and other documents:

Class 3600	1998
Class 3611	1999
Class 3810	1989

Original Project ID: 3016395

FM Approval Granted: August 27, 2003

Subsequent Revision Reports / Date FM Approval Amended

Report Number	Date	Report Number	Date
---------------	------	---------------	------

FM Global Technologies LLC

David W. Styrula  
Technical Team Manager

8/28/03  
Date

An FM Global Enterprise