



CERTIFICATION RECORD

The company named below has been authorized by CSA International to represent the products listed in this record as "CSA Certified" and to affix the CSA Mark to these products according to the terms and conditions of the CSA Service Agreement and applicable CSA program requirements (including additional Markings).

File No: 108175_0_000

Class No: 4812 06 SIGNAL APPLIANCES Miscellaneous

SUBMITTOR

4529993 Teledyne Instruments
Analytical Instruments
16830 Chestnut St.
City of Industry, CA 91749-1020
USA

FACTORIES

4529993 Teledyne Instruments
Analytical Instruments
16830 Chestnut St.
City of Industry, CA 91749-1020
USA

October 7, 1996

- Oxygen analyzer, Model 3290, 100-240V, 50/60Hz, 0.3A max; alarm relays, 125V ac, 0.5A (non-ind).

Note: Certified as a component for use in other equipment where the suitability of the combination is to be determined in the final application.