



FM Approvals
1151 Boston Providence Turnpike
P.O. Box 9102 Norwood, MA 02062 USA
T: 781 762 4300 F: 781-762-9375 www.fmapprovals.com

Member of the FM Global Group

CERTIFICATE OF COMPLIANCE

HAZARDOUS (CLASSIFIED) LOCATION ELECTRICAL EQUIPMENT

This certificate is issued for the following equipment:

3110 Oxygen Analyzer
IS//I/1/ABCD/T6 Ta = 40 °C; A82919

Equipment Ratings:

Intrinsically Safe for use in Class I, Division 1, Groups A, B, C, D Hazardous (Classified) Locations.

FM Approved for:

Teledyne Analytical Instruments
16830 Chestnut St
City of Industry, CA 91749



This certifies that the equipment described has been found to comply with the following Approval Standards and other documents:

3600	1998
3610	2007
3810	2005

Original Project ID: 3032537

Approval Granted: *August 11, 2009*

Subsequent Revision Reports / Date Approval Amended

Report Number	Date	Report Number	Date
---------------	------	---------------	------

FM Approvals LLC

J. E. Marquedant

J. E. Marquedant
Group Manager, Electrical

11 August 2009

Date