



FM Approvals
1151 Boston Providence Turnpike
P.O. Box 9102 Norwood, MA 02062 USA
T: 781 762 4300 F: 781-762-9375 www.fmapprovals.com

Member of the FM Global Group

CERTIFICATE OF COMPLIANCE

HAZARDOUS LOCATION ELECTRICAL EQUIPMENT PER CANADIAN REQUIREMENTS

This certificate is issued for the following equipment:

3110 Oxygen Analyzer
IS/II/1/ABCD/T6 Ta = 40 °C; A82919

Equipment Ratings:

Intrinsically Safe for use in Class I, Division 1, Groups A, B, C, D Hazardous (Classified) Locations.

FM Approved for:

Teledyne Analytical Instruments
16830 Chestnut St
City of Industry, CA 91749



Member of the FM Global Group

This certifies that the equipment described has been found to comply with the following Approval Standards and other documents:

CSA C22.2 No 0 2006
CSA C22.2 No 157 2006

Original Project ID: 3032537

Approval Granted: *August 11, 2009*

Subsequent Revision Reports / Date Approval Amended

Report Number	Date	Report Number	Date
---------------	------	---------------	------

FM Approvals LLC

J. E. Marquedant

J. E. Marquedant
Group Manager, Electrical

11 August 2009
Date