

Certificate of Compliance

Certificate Number 101405-E257873
Report Reference E257873-February 15, 2000
Issue Date 2005 October 14

Page 1 of 1



**Underwriters
Laboratories Inc.**

Issued to: **TELEDYNE INSTRUMENTS ANALYTICAL**
16830 CHESTNUT ST
CITY OF INDUSTRY, CA 91749

*This is to certify that
representative samples of*

TELEMETERING EQUIPMENT


8800A, 8800B and 8800NB, provide intrinsically safe outputs when installed in accordance with control drawing XDT 00.D.6003 for use in Class I, Groups A, B, C and D; Class II, Groups E, F and G hazardous locations.

*Have been investigated by Underwriters Laboratories Inc.® in
accordance with the Standard(s) indicated on this Certificate.*

Standard(s) for Safety: **UL 913**

Additional Information: **N/A**

Only those products bearing the UL Listing Mark should be considered as being covered by UL's Listing and Follow-Up Service.

The UL Listing Mark generally includes the following elements: the symbol UL in a circle:  with the word "LISTED"; a control number (may be alphanumeric) assigned by UL; and the product category name (product identifier) as indicated in the appropriate UL Directory.

Look for the UL Listing Mark on the product

Issued by:

CHRISTOPHER REES
Underwriters Laboratories Inc.

Reviewed by:

BEN SCHAEFFER
Underwriters Laboratories Inc.