



# Certificate of Compliance

**Certificate:** 2671602

**Master Contract:** 259525

**Project:** 2671602

**Date Issued:** December 12, 2013

**Issued to:** Teledyne Instruments  
Analytical Instruments  
16830 Chestnut St.  
City of Industry, CA 91749-1020  
USA

**Attention:** Mr. Roger Starlin

*The products listed below are eligible to bear the CSA Mark shown*



*Kim Atkins*

**Issued by:** K. Atkins

## PRODUCTS

CLASS 2252 05 - PROCESS CONTROL EQUIPMENT

Dewpoint Transmitters for use in Non-Hazardous Locations, permanently connected, rated: 100-240V ac, 10VA, 50/60Hz, ambient 50°C, Pollution degree 2, Installation Category II. Output contact relay: 10A-30V dc, 240V ac. Model 8800A, panel mount enclosure.

## APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements



**Certificate:** 2671602

**Master Contract:** 259525

**Project:** 2671602

**Date Issued:** December 12, 2013

---

## **MARKINGS**


The manufacturer is required to apply the following markings:

- Products shall be marked with the markings specified by the particular product standard.
- Products certified for Canada shall have all Caution and Warning markings in both English and French.

Additional bilingual markings not covered by the product standard(s) may be required by the Authorities Having Jurisdiction. It is the responsibility of the manufacturer to provide and apply these additional markings, where applicable, in accordance with the requirements of those authorities.

General: Markings appear as outlined in the Certification Report on an adhesive nameplate, on a screw or rivet secured metal nameplate or in some other permanent manner. Labels, if shown in the report, may vary slightly in layout, provided that the required information continues to be shown.

The following markings appear on the product:

1. Company's identification (company name and/or file number and/or registered tradename).
2. Marking on the unit that indicates the manufacturing location if the equipment is manufactured at more than one factory location.
3. Model designation.
4. Electrical rating (input voltage, current or volt amperes).
5. Date of manufacture: Month and year of manufacture or date code. If a serial number is used instead of date of manufacture, a record of serial numbers shall be kept traceable to date of manufacture. (Not related to date of sale).
6. The CSA Monogram .