



# Certificate of Compliance

**Certificate:** 1390739 (LR 68348)

**Master Contract:** 163296

**Project:** 2064525

**Date Issued:** 2008/07/11

**Issued to:** Teledyne Instruments  
Analytical Instruments  
16830 Chestnut St.  
City of Industry, CA 91749-1020  
USA  
Attention: Mr. Roger Starlin

*The products listed below are eligible to bear the CSA Mark shown*



**Issued by:** Cindy Johnson, ASCT.

**Authorized by:** Ken Rutledge, Product Group  
Manager

## **PRODUCTS**

**CLASS 8750 01 - MEDICAL ELECTRICAL EQUIPMENT**

Oxygen Monitors, Models MX300, MX300-I, AX300, AX300-I battery operated, no patient applied parts.

## **APPLICABLE REQUIREMENTS**

CAN/CSA C22.2 No 601.1-M90 - Safety of Medical Electrical Equipment, Part I, General Requirements for Safety

CSA 601.1 Supplement 1:1994



CSA INTERNATIONAL

**Certificate:** 1390739 (LR 68348)

**Master Contract:** 163296

**Project:** 2064525

**Date Issued:** 2008/07/11

---

CSA 601.1 Amendment 2:1998