



Member of the FM Global Group

FM Approvals
1151 Boston Providence Turnpike
P.O. Box 9102 Norwood, MA 02062 USA
T: 781 762 4300 F: 781-762-9375 www.fmapprovals.com

CERTIFICATE OF COMPLIANCE

HAZARDOUS (CLASSIFIED) LOCATION ELECTRICAL EQUIPMENT

This certificate is issued for the following equipment:

Insta Trans-XD abcde. Trace and Percent Oxygen Digital Transmitter.

IS/I/1/ABCD/T4 Ta = 0°C to +50°C – D-92719; Entity;

NI/I/2/ABCD/T4 Ta = 0°C to +50°C – D-92719; NIFW;

Entity Parameters:

Ui = 28Vdc, Ii = 93mA, Pi = 0.66W, Ci = 1.1nF, Li = 0

Nonincendive Field Wiring Parameters:

Ui = 28Vdc, Ii = 27mA, Pi = 0.66W, Ci = 1.1nF, Li = 0

a = Sensor: A (Std.) or B (A5).

b = Cell Holder: A (Std. B2/L2C Profile) or B (Insta-Trace).

c = Retaining Ring: A (Std. UHMW ply) or B (316 SS).

d = Gas Connection: A (1/8" NPTF), B (Std. 1/8" T), C (1/4" T) or D (6mm).

e = O Ring: A (Std. Viton), B (Kalrez), C (EPR), D (Neoprene) or E (Bun-n).

Special Conditions of Use:

- 1. The enclosure contains plastic. To prevent the risk of electrostatic sparking the plastic surface should be cleaned only with a damp cloth.*
- 2. The apparatus enclosure contains aluminum and is considered to constitute a potential risk of ignition by impact or friction. Care must be taken into account during installation and use to prevent impact or friction.*

Equipment Ratings:

Intrinsically Safe for Class I, Division 1, Groups A, B, C and D in accordance with drawing D-92719;

Nonincendive with Nonincendive Field Wiring for Class I, Division 2, Groups A, B, C and D in accordance with drawing D-92719 hazardous (Classified) locations Temperature Class T4 Ta = 0°C to +50°C.

FM Approved for:

Teledyne Analytical Instruments
City Of Industry, CA USA



Member of the FM Global Group

This certifies that the equipment described has been found to comply with the following Approval Standards and other documents:

Class 3600	2011
Class 3610	2010
Class 3611	2004
Class 3810	2005

Original Project ID: 3052658

Approval Granted: October 25, 2015

Subsequent Revision Reports / Date Approval Amended

Report Number	Date	Report Number	Date
RR203058	November 10, 2015		

FM Approvals LLC

J.E. Marquedant
Manager, Electrical Systems

10 November 2015

Date