



Member of the FM Global Group

FM Approvals
1151 Boston-Providence Turnpike
P.O. Box 9102 Norwood, MA 02062 USA
T: 781 762 4300 F: 781 762 9375 www.fmapprovals.com

CERTIFICATE OF COMPLIANCE

HAZARDOUS LOCATION ELECTRICAL EQUIPMENT PER CANADIAN REQUIREMENTS

This certificate is issued for the following equipment:

3001. Insta Trans Oxygen Transmitter.
IS / I / 1 / ABCD / T4 Ta = 50°C — D-73226; NI / I / 2 / ABCD / T4 Ta = 50°C

Equipment Ratings:

Intrinsically safe, for use in Class I, Division 1, Groups A, B, C, D, in accordance with manufacturer's Control Drawing No. D-73226; Nonincendive for Class I, Division 2, Groups A, B, C, D Hazardous (Classified) Locations

FM Approved for:

Teledyne Instruments-Analytical Instruments
16830 Chestnut Street
City of Industry, CA 91749

This certifies that the equipment described has been found to comply with the following Approval Standards and other documents:

CSA-C22.2 No. 157	1992
CSA-C22.2 No. 213	1987
CSA-C22.2 No. 142	1987

Original Project ID: 3007225


Canadian Project ID: 3027270

Approval Granted: *JUNE 12, 2006*

Subsequent Revision Reports / Date Approval Amended

Report Number	Date	Report Number	Date
---------------	------	---------------	------

FM Approvals LLC



Roger L. Allard
Assistant Vice President

JUNE 12, 2006
Date