

DECLARATION OF CONFORMITY

APPLICATION OF COUNCIL DIRECTIVE : 94/9/EC

STANDARDS TO WHICH CONFORMITY IS DECLARED : EN 60079-0 : 2006 (IEC 60079-0 : 2004)
EN 60079-11 : 2007 (IEC 60079-11 : 2006)
EN 60079-26 : 2004 (IEC 60079-26 : 2004)

MANUFACTURER'S NAME : TELEDYNE ANALYTICAL INSTRUMENTS


MANUFACTURER'S ADDRESS : 16830 Chestnut Street
City of Industry, CA 91748
U.S.A.

TYPE OF EQUIPMENT : Portable Oxygen Analyzer 3110

EQUIPMENT CLASS : Electrical Equipment Measurement, Control & Laboratory Use -Industrial

MODEL NUMBER : 3110

I, THE UNDERSIGNED, HEREBY DECLARE THAT THE EQUIPMENT SPECIFIED ABOVE CONFORMS TO THE ABOVE STANDARD(S) PER 94/9/EC and have been type-approved by ISSep, Zoning A. Schweitzer, rue de la Platinerie, B-7340 Colfontaine. ISSep notified body identification number: 0492. Mark: Ex II 2 G. Code: Ex ia II C T3

SIGNATURE: 

FULL NAME: Roger Starlin

POSITION: QA Manager

Date: 1-22-09

PLACE: City of Industry, California

SIGNATURE: 

FULL NAME: Stephen Broy

POSITION: Director of Engineering

Date: 1-22-09

PLACE: City of Industry, California