

# EU DECLARATION OF CONFORMITY

---

**APPLICATION OF COUNCIL DIRECTIVE** : 2014/34/EU

**STANDARDS TO WHICH CONFORMITY IS DECLARED** : EN 60079-0 : 2012+A11:2014  
(IEC 60079-0 : 2011)  
EN 60079-2 : 2014  
(IEC 60079-2 : 2014-07)

**MANUFACTURER'S NAME** : TELEDYNE ANALYTICAL INSTRUMENTS

**MANUFACTURER'S ADDRESS** : 16830 Chestnut Street  
City of Industry, CA 91748  
U.S.A.

**TYPE OF EQUIPMENT** : Hydrocarbon-in-Water Analyzer System

**EQUIPMENT CLASS** : Ex II 2 G Ex pxb IIC T\* Gb  
(\*See model description for T-Class details pg. 2)

**MODEL NUMBER** : 4000 Series

---

I, THE UNDERSIGNED, HEREBY DECLARE THAT THE EQUIPMENT SPECIFIED ABOVE CONFORMS TO THE ABOVE STANDARD(S) PER 2014/34/EU and have been type-approved by Intertek Testing & Certification Limited, Intertek House, Cleeve Road, Leatherhead, Surrey, KT22 7SB. Intertek notified body identification number: 0359. Mark: Ex II 2 G. Code: Ex pxb IIC T\* Gb. EC Type Certificate Number: ITS14ATEX181526X  
IECEx Certificate Number: IECEx ITS 14.0065X


SIGNATURE: 

FULL NAME: Roger Starlin

POSITION: QA Manager

Date: 2-2-2017

PLACE: City of Industry, California

SIGNATURE: 

FULL NAME: Angel Alegria

POSITION: New Products Manager

Date: 2/02/17

PLACE: City of Industry, California

# EU DECLARATION OF CONFORMITY

---

**MODEL NUMBER : 4000 Series**

**Models covered by this Declaration of Conformity are:**

<u>Model</u>	<u>T Code</u>
• 4080/4020B	T3
• 4080/4030B	T3
• 4080/4040B	T3
• 4080/4060B	T3
• 4020B	T4
• 4030B	T4
• 4040B	T4
• 4060B	T4