

EU DECLARATION OF CONFORMITY

APPLICATION OF COUNCIL DIRECTIVE : 2014/34/EU

STANDARDS TO WHICH CONFORMITY IS DECLARED : EN 60079-0 : 2012+A11:2014
EN 60079-2 : 2014

MANUFACTURER'S NAME : TELEDYNE ANALYTICAL INSTRUMENTS

MANUFACTURER'S ADDRESS : 16830 Chestnut Street
City of Industry, CA 91748
U.S.A.

TYPE OF EQUIPMENT : Process Analyzer System

EQUIPMENT CLASS : Ex II 2 G Ex ib pxb IIC T4. Gb

MODEL NUMBER : 6700

I, THE UNDERSIGNED, HEREBY DECLARE THAT THE EQUIPMENT SPECIFIED ABOVE CONFORMS TO THE ABOVE STANDARD(S) PER 2014/34/EU and have been type-approved by Intertek Testing & Certification Limited, Intertek House, Cleeve Road, Leatherhead, Surrey, KT22 7SB. Intertek notified body identification number: 0359. Mark: Ex II 2 G. Code: Ex ib pxb IIC T4. Gb.

EC Type Certificate Number: ITS15ATEX18336X

SIGNATURE: 

FULL NAME: Roger Starlin

POSITION: QA Manager

Date: 2-2-2017

PLACE: City of Industry, California

SIGNATURE: 

FULL NAME: Angel Alegria

POSITION: New Products Manager

Date: 2/02/17

PLACE: City of Industry, California