

# EU DECLARATION OF CONFORMITY

This declaration of conformity is issued under the sole responsibility of the manufacturer

---

APPLICATION OF COUNCIL : 2014/34/EU  
ATEX DIRECTIVE

STANDARDS TO WHICH : EN 60079-0:2012  
CONFORMITY IS DECLARED : EN 60079-11:2012

MANUFACTURER'S NAME : Teledyne Analytical Instruments

MANUFACTURER'S ADDRESS : 16830 Chestnut Street  
City of Industry, CA 91748  
U.S.A.

TYPE OF EQUIPMENT : Oxygen Measuring Transmitter  
Type Insta Trans

EQUIPMENT CLASS : Ex II 2 (1) G Ex ia IIC T4 Gb  
(0°C ≤ Ta ≤ +50°C)

MODEL NUMBER : 3001 Insta Trans

---

I, the undersigned, hereby declare that the equipment specified above conforms to the referenced standards (s) per 2014/34/EU and have been type-approved by SGS Baseefa Limited, Staden Lane, Buxton, Derbyshire SK17 9RZ, UK. Baseefa notified body identification number: 1180. Mark: Ex II 2 (1) G. Code: Ex ia IIC T4 Gb (0°C ≤ Ta ≤ + 50°C). EC Type Certificate Number: BAS01ATEX2200X


SIGNATURE: 

FULL NAME: Roger Starlin

POSITION: QA Manager

Date: 7-27-18

PLACE: City of Industry, California

SIGNATURE: 

FULL NAME: Angel Alegria

POSITION: New Products Manager

Date: 7-27-18

PLACE: City of Industry, California