

EU DECLARATION OF CONFORMITY

APPLICATION OF COUNCIL ATEX DIRECTIVE : 2014/34/EU

STANDARDS TO WHICH CONFORMITY IS DECLARED : EN 60079-0:2012
EN 60079-11:2012

MANUFACTURER'S NAME : Teledyne Analytical Instruments

MANUFACTURER'S ADDRESS : 16830 Chestnut Street
City of Industry, CA 91748
U.S.A.

TYPE OF EQUIPMENT : Oxygen Measuring Transmitter
Type Insta Trans

EQUIPMENT CLASS : Ex II 2 (1) G Ex ia IIC T4 Gb
(0°C ≤ Ta ≤ +50°C)

MODEL NUMBER : 3001 Insta Trans

I, the undersigned, hereby declare that the equipment specified above conforms to the referenced standards (s) per 2014/34/EU and have been type-approved by SGS Baseefa Limited, Staden Lane, Buxton, Derbyshire SK17 9RZ, UK. Baseefa notified body identification number: 1180. Mark: Ex II 2 (1) G. Code: Ex ia IIC T4 Gb (0°C ≤ Ta ≤ + 50°C). EC Type Certificate Number: BAS01ATEX2200X


SIGNATURE: 

FULL NAME: Roger Starlin

POSITION: QA Manager

Date: 2-2-2017

PLACE: City of Industry, California

SIGNATURE: 

FULL NAME: Angel Alegria

POSITION: New Products Manager

Date: 2/02/17

PLACE: City of Industry, California