

DECLARATION OF CONFORMITY

APPLICATION OF COUNCIL DIRECTIVE : 89/336/EEC

STANDARDS TO WHICH CONFORMITY IS DECLARED : EN55011 Class A Group I
EN61000-4-2 EN61000-4-5
EN61000-4-3 EN61000-4-6
EN61000-4-4 EN61000-4-11

MANUFACTURER'S NAME : TELEDYNE ANALYTICAL INSTRUMENTS

MANUFACTURER'S ADDRESS : 16830 Chestnut Street
City of Industry, CA 91748-1020
U.S.A.

TYPE OF EQUIPMENT : Helium Oxygen Analyzer

EQUIPMENT CLASS : Laboratory, Measurement & Process Control
Equipment: Industrial Environment

MODEL NUMBER : Mixchek Models
Wall mount Helium = MCWHE
Wall mount Helium/O2 = MCWHEO2
Portable Helium = MCPHE
Portable Helium/O2 = MCPHEO2

I, THE UNDERSIGNED, HEREBY DECLARE THAT THE EQUIPMENT SPECIFIED ABOVE CONFORMS TO THE ABOVE DIRECTIVE(S) AND STANDARD(S).

SIGNATURE: 

FULL NAME: Steve Broy

POSITION: Director of Engineering

DATE: 5/27/03

PLACE: City of Industry, California