



- ⑥ FOR UNITS WITH RANGE OF 0-1 PPM OR LESS SEE SCHEMATIC B-11213
- ⑤ DELETE Z1 FOR 0-200 PPM O₂ RANGE UNITS ONLY! ADJUST: R11 TO 0 WITH R7 FCCW; R3 TO 8 mA DC R7 FCCW
- ④ SEE DWG# B-17969 FOR WIRING DIAGRAM OF VARIOUS INPUT VOLTAGES.
- ③ RESISTOR VALUES SELECTED PER APPLICATION.
2. ALL CAPACITORS ARE IN MICROFARADS.
1. ALL RESISTORS ARE IN OHMS, METAL FILM 1/8W ±1%.
UNLESS OTHERWISE SPECIFIED
GENERAL NOTES

THIS DRAWING IS THE PROPERTY OF TELETYPE ANALYTICAL INSTRUMENTS AND CONTAINS CONFIDENTIAL INFORMATION. IT IS NOT TO BE COPIED, REPRODUCED OR USED WITHOUT WRITTEN PERMISSION.

TOLERANCE UNLESS OTHERWISE SPECIFIED	FRACT. ± 1/64	DEC. ± .005	ANGULAR ± 1/2°
ANALYTICAL INSTRUMENTS		SCALE	DATE 10-1-86
A TELETYPE COMPANY SAN GABRIEL, CALIFORNIA		MATL	
MDL 306 WAX SCHEMATIC DIAG. (220V)		FINISH	
DR. L.M.	ENGR.	APP.	
D-41307			

S.D. 131874

SIMILAR TO: D19081