

SIMDIS Analysis for Gasoline Blending

Purpose

Lab analysis for product specification and certification and online blending control

Measured Components

- IBP, 10%, 20%, 30%, 40%, 50%, 60%, 70%, 80%, 90%, 95.5% and FBP analysis for the determination of the boiling range distribution of gasoline and gasoline components.
- Applicable to petroleum products and fractions with a final boiling point of 500°F (260°C) or lower.

Cycle Time

< 300 seconds (< 5 minutes)

Detector(S)

FID

Reference Methods

ASTM 3710, ASTM 7096

*Contact Teledyne for specifications for this application.



16830 Chestnut Street, City of Industry, CA 91748 USA
Phone: 626-934-1500/888-789-8168 Fax: 626-934-1651
EMAIL: Ask_tai@teledyne.com www.teledyne-ai.com

Contact the TAI Sales Team to ask about our other instruments.