DECLARATION OF CONFORMITY

MOCA 18 (18 MICE)	
APPLICATION OF COUNCIL	: 2014/30/EU
STANDARDS TO WHICH CONFORMITY IS DECLARED:	EN61326-1: 2013 EN55011 Class A Group 1
	EN61000-4-2, EN61000-4-3, EN610004-4 EN61000-4-5, EN61000-4-6, EN61000-4- EN61000-4-11
	EN61010-1: 2010 LVD
MANUFACTURER'S NAME	: TELEDYNE ANALYTICAL INSTRUMENTS
MANUFACTURER'S ADDRESS	: 16830 Chestnut Street City of Industry, CA 91748 U.S.A.
TYPE OF EQUIPMENT	: Oxygen Analyzer
EQUIPMENT CLASS	: Electrical Equipment Measurement, Control & Laboratory Use- Industrial
MODEL NUMBER	: 3000TAXL-EU
	Y DECLARE THAT THE EQUIPMENT S TO THE ABOVE DIRECTIVE(S) AND
	SIGNATURE: Alegria
	FULL NAME: Angel Alegria
	POSITION: New Products Manager
4	DATE: 4 - 1/ - 17
	PLACE: City of Industry, CA