

Model 3010PBC

Orders should include the part number, model number, and instrument serial number. Call Teledyne Analytical Instruments at +1-888-789-8168 (USA Only) or +1-626-934-1500 to place an order for spare parts. Contact your regional sales representative for instrumentation questions; see [http://ai.www-
uat.teledyne.com/en-us/who-we-are/worldwide-representatives](http://ai.www-uat.teledyne.com/en-us/who-we-are/worldwide-representatives).

Recommended Spares		
QTY	Part No.	Description
1	D-65295B	Customer interface board, control unit
1	C-62371B	Front panel board
1	C-62368B	Percent preamplifier board
1*	C-73870C	Main computer board (standard)
1*	C-73870A	Main computer board (4-20 mA option)
1	C-65407A	Interface board
2	F1275	Fuse, 1A, 250V, 5x20 mm, T-Slowblow (EU)
2	O38	O-ring
1**	C6689-B1	Micro-fuel cell, class B-1
1	I226	Isolated 4-20 mA module
* Order C or A as appropriate ** C6689-B1 is used on the standard percent model		



16830 Chestnut Street, City of Industry, CA 91748 USA
Phone: 626-934-1500/888-789-8168 Fax: 626-934-1651
EMAIL: Ask_tai@teledyne.com www.teledyne-ai.com

Find your local sales representative: [http://ai.www-
uat.teledyne.com/en-us/who-we-are/worldwide-representatives](http://ai.www-uat.teledyne.com/en-us/who-we-are/worldwide-representatives)