

Model 3010TAC

Orders should include the part number, model number, and instrument serial number. Call Teledyne Analytical Instruments at +1-888-789-8168 (USA Only) or +1-626-934-1500 to place an order for spare parts. Contact your regional sales representative for instrumentation questions; see [http://ai.www-
uat.teledyne.com/en-us/who-we-are/worldwide-representatives](http://ai.www-uat.teledyne.com/en-us/who-we-are/worldwide-representatives).

Recommended Spares		
QTY	Part No.	Description
1	C-67821A	Back panel board
1	C-62371	Front panel board
1	C-62368A	Trace preamplifier board
1	C-73870C	Main computer board (standard)
5	F9	Fuse, 1A, 250V, 3AG, Slowblow (US)
5	F1275	Fuse, 1A, 250V, 5x20 mm, T-Slowblow (EU)
1	O-165	O-ring
1	C6689-L2C	Micro-fuel cell, class L-2C
2	B367	Intrinsic safety Zener barrier
1	A-68314	Back panel connector kit
1	I226	Isolated 4-20 mA module
1	B-66378	Cell spacer for B-2C or A-2C sensor



16830 Chestnut Street, City of Industry, CA 91748 USA
Phone: 626-934-1500/888-789-8168 Fax: 626-934-1651
EMAIL: Ask_tai@teledyne.com www.teledyne-ai.com

Find your local sales representative: [http://ai.www-
uat.teledyne.com/en-us/who-we-are/worldwide-representatives](http://ai.www-uat.teledyne.com/en-us/who-we-are/worldwide-representatives)